

BLANK PAGE



Indian Standard

SPECIFICATION FOR INTERMITTENT POSITIVE PRESSURE RESPIRATOR, BAG TYPE, MANUALLY OPERATED

(Second Reprint OCTOBER 1993)

UDC 614.894.4:615.835.32

© Copyright 1971

BUREAU OF INDIAN STANDARD S MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG NEW DELHI 110002

Gr 1 November 1971

Indian Standard

SPECIFICATION FOR INTERMITTENT POSITIVE PRESSURE RESPIRATOR. BAG TYPE. MANUALLY OPERATED

Anaesthesia, Resuscitation and Allied Equipment Sectional Committee, CPDC 13

Rebresenting Chairman

DRB. L. BHATTACHARIEE Directorate General of Health Services, New Delhi

Members

SHRIBALESHWAR SAHAI BHARCAVA Khushwagat Industries, Rewari (Haryana)

SHRI SATISHWAR SAHAI BHARGAVA (Alternate)

Institute of Post Graduate Medical Education & DR A. K. BOSE Research, Calcutta

SHRI H. GHOSH Indian Oxygen Ltd, Calcutta

SHRI A. N. MUKERJI (Alternate)

SHRI V. KRISHNAMOORTHY Directorate General of Technical Development, New Delhi

SHRI H. C. MUKHERJEE Ministry of Defence (DGI)

Dr V. Rajagopalan

Health and Family Planning Department, Government of Tamil Nadu, Madras

Military Hospital [Ministry of Defence (DGAFMS)], New Delhi BRIG K. R. RAMA RAO Defence

Jawaharlal Institute of Post Graduate Medical DR N. P. SINGH

Education & Research, Pondicherry

Director General, ISI (Ex-officio Member) SHRI A. B. RAO.

Director (Consr Prod) Secretary

SHRI D. K. AGARWAL

Assistant Director (Consr Prod), IS1

Anaesthesia and Oxygen Therapy Equipment Subcommittee, CPDC 13:1

Convener

SHRI H. GHOSH Indian Oxygen Ltd, Calcutta

Members

SHRI A. N. MUKERJI (Alternate to

Shri H. Ghosh)

SHRI BALESHWAR SAHAI BHARCAVA Khushwagat Industries, Rewari (Haryana)

SHRI P. S. BHARGAVA (Alternate)

Directorate General of-Health Services, New Delhi

Directorate General of-Health Services, New Delhi

Directorate General of-Health Services, New Delhi DRB. L. BHATTACHARJEE cary Hospital [Ministry of (DGA-FMS) **],** New Delhi BRIGK R. RAMA RAO Military Defence

DRS. S. VERMA Railway Board (Ministry of Railways)

DR K. B. SUNDERAM (Alternate)

SHRI D. N. VIG SHRI R. N. VIG (Alternate)

Ultradent Private Ltd, Bombay

BUREAU OF INDIAN STANDARDS MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG NEW DELHI 110002

Indian Standard

SPECIFICATION FOR INTERMITTENT POSITIVE PRESSURE RESPIRATOR, BAG TYPE. MANUALLY OPERATED

O. FOREWORD

- 0.1 'This Indian Standard was adopted by the Indian Standards Institution on 15 June 1971, after the draft finalized by the Anaesthesia, Resuscitation and Allied Equipment Sectional Committee had been approved by the Consumer Products Division Council.
- 0.2 This standard has been formulated at the instance of the Advisory Committee for Development of Surgical Instruments, Equipment and Appliances, Government of India.
- 0.3 This standard is one of a series of Indian Standards on diagnostic and anaesthetic equipment.
- 0.4 This standard contains clauses which call for agreement between the purchaser and the supplier. The relevant clauses are 4.1 and 6.1.
- 0.5 **For the** purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test or analysis, shall be rounded off in accordance with IS:2-1960*. The number of significant places retained in the **rounded off** value should be the same as that of the specified value in this standard.

1. SCOPE

1.1 This standard prescribes the requirements for manually operated bag type intermittent positive pressure respirator used for artificial respiration and resuscitation.

2. MATERIALS

2.1 Bag — The hag shall be made of good quality natural ruhher which when tested for ageing in an air-oven for 168 hours at 70 \pm 1°C does not show appreciable stiffening, softening and cracking or any other

^{*}Rules for rounding off numerical values (revised).

change in condition. The rubber used shall have minimum tensile strengths of 105 and 95 kgf/cm^2 (10 and 9.3 MN/m^2) and minimum elongations at break of 400 and 300 percent before and after ageing respectively. Suitable plastics or polymer materials may also be used.

2.2 Connection Tube -Any suitable material may be used for the connection tube.

3. WORKMANSHIP AND FINISH

3.1 The bag shall be adequately reinforced at the neck, and the wall of the bag except for the reinforced areas shall be of uniform thickness. The surfaces of the bag shall be smooth and free from pinholes, wrinkles, creases, embedded foreign matter and other defects. All the metallic parts shall be plated chromium over nickel in accordance with Service Grade No. 2 of IS:1068-1968*.

4. REQUIREMENTS

- 4.1 Rubber face masks shall in every respect conform to IS:6190-1971†. Alternatively, face masks made of suitable transparent material and satisfying the requirements of 5 and 8 of IS:6190-1971 † may be used.
- 4.2 The respirator shall be fitted with a non-rebreathing valve made of sturdy and unbreakable clear plastics material. When the bag is pressed, air shall flow to the patient through the valve and the face mask. When the bag is released, the valve shall open towards outside and expiration shall take place with least resistance. During the expiration phase, air shall automatically fill in the bag. The valve at any stage shall not stick.
- 4.3 The face mask end of the valve shall have male conical fittings of the adult size conforming to IS:5413-1969‡.
- 4.4 The inlet shall be provided for connection to an oxygen source. The tube shall have an additional arrangement for preventing air to **escape** into the atmosphere when the bag is squeezed.
- 4.5 Outer size of the bag shall be such that it can be grasped **easily** with one hand. The bag shall be self-inflating and shall resume its original shape after each depression. The bag shall be smooth and shall not hurt the finger tips. The operation of the respirator shall be effortless and the hand fatigue shall be negligible.

^{*}Specification for electroplated coatings of nickel and -chromium on iron and steel $(first\ revision\)$.

[†]Specification for anaesthetic face masks.

[‡]Specification for conical fittings for to-and-fro carbon dioxide absorber (Waters type) for use with gas anaesthetic apparatus.

IS: 6194 - 1971

- **4.6** The axial length of the neck of the bag shall be not less than 50 mm. The thickness of the rubber forming the neck of the bag shall be not less than 2.5 mm.
- 4.7 When-the opening of the bag is closed securely, the bag immersed in **water and** pressed, there shall be no sign of escape of air through any part of the bag.
- 4.8 The bag shall be capable of withstanding exposure to saturated steam at a temperature of 134°C corresponding to a gauge pressure of 2.2 kgf/cm² for not less than 20 minutes in a sterilizer without loss of shape or other significant deterioration.
- 4.9 When manually operated, the respirator shall deliver a tidal air of volume 1 200 to 1 600 ml.

5. MARKING

- **5.1** The respirator shall be marked with the manufacturer's name, initials or trade-mark on the bag.
 - 5.1.1 The respirator may also be marked with the Standard Mark.

NOTE — The use of the Standard Mark is governed by the provisions of the Bureau of Indian Standards Act, 1986 and the Rules and Regulations made thereunder. The Standard Mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard under a well defined system of inspection, testing and quality control which is devised and supervised by BIS and operated by the producer. Standard marked products are also continuously checked by BIS for conformity to that standard as a further safeguard. Details of conditions under which a licence for the use of the Standard Mark may be granted to manufacturers or producers may be obtained from the Bureau of Indian Standards.

6. PACKING

6.1 The packing shall be done as agreed to between the purchaser and the supplier. It is however recommended that the various parts of the respirator should be dismantled and packed in a strong carrying case.

BUREAU OF -INDIAN STANDARDS

Headquarters:

Manak Bhavan, 9 Bahadur Shah Zafar Marg, NEW DELHI	110002
Telephones: 331 01 31, 331 13 75 Telegrams: M (Common to	anaksanstha
Regional Offices:	Telephone
Central : Manak Bhavan, 9 Bahadur Shah Zafar Marg, NEW DELHI 110002	331 01 31 331 1375
*Eastern :1/14 C. I. T. Scheme VII M, V. I. P. Road, Maniktola, CALCUTTA 700054	36 24 99
Northern: SCO 445-446, Sector 35-C, CHANDIGARH 160036	21843 ₁ 3 16 41
Southern: C. I. T. Campus, MADRAS 600113	41 24 42 41 25 19 41 29 16
†Western: Manakalaya, E9 MIDC, Marol, Andheri (East), BOMBAY 400093	6 32 92 95
Branch Offices:	
'Pushpak', Nurmohamed Shaikh Marg, Khanpur, AHMADABAD 380001	2 63 48 1 2 63 49
‡Peenya Industrial Area 1st Stage, Bangalore Tumkur Road BANGALORE 560058	38 49 55 38 49 56
Gangotri Complex, 5th Floor, Bhadbhada Road, T. T. Nagar, BHOPAL 462003	667 16
Plot No. 82/83, 'Lewis Road, BHUBANESHWAR 751002 53/5, Ward No. 29, R.G. Barua Road, 5th Byelane, GUWAHATI 781003	5 36 27 3 31 77
5-8-56C L. N. Gupta Marg (Nampally Station Road), HYDERABAD 500001	23 1083
R14 Yudhister Marg, C Scheme, JAIPUR 302005	$ \left\{ \begin{array}{c} 6 \ 34 \ 71 \\ 6 \ 98 \ 32 \end{array} \right. $
117/418 B Sarvodaya Nagar, KANPUR 208005	21 68 76 21 82 92
Patliputra Industrial Estate, PATNA 800013 T.C. No. 14/1421. University P.O Palayam	6 23 05
TRIVANDRUM 695035	16 21 17
Inspection Offices (With Sale Point):	
Pushpanjali, First Floor, 205-A West High Court Road, Shankar Nagar Square. NAGPUR 440010	25171
Institution of Engineers (India) Building, 1332 Shivaji Nag PUNE 411005	ar, 5 24 35
*Sales Office in Calcutta is at 5 Chowringhee Approach, P. 0. Pri	ncep 27 68 00

^{*}Sales Office in Calcutta is at 5 Chowringhee Approach, P. 0. Princep 27 68 00 Street, Calcutta 700072

[†]Sales Office in Bombay is at Novelty Chambers, Grant Road, 89 66 28 Bombay 400007

[‡]Sales Office in Bangalore is at Unity Building, Narasimharajs Square, 22 36 71 Bangalore 560002